

Reference Device for Calibration of Radon Exhalation Rate Measuring Instruments and Its Performance

Authors: Li Weigang, Xiao Detao, He Zhengzhong, Deng Xiangyuan, Qiu Shoukang, Xiao Detao

Date: 2023-07-07T00:00:00+00:00

Abstract

Environmental radon originates from the exhalation and release of soil, rocks, and building materials. Environmental radon contamination tracing and radon pollution prevention and control require measurement of the radon exhalation rate on media surfaces. Reliable measurements of the radon exhalation rate cannot be achieved without regular calibration of the measuring instrument using a high-performance reference device. In this study, a reference device for calibrating radon exhalation rate measuring instruments was developed, incorporating a diffusion solid radon source with a high and stable radon emanation coefficient, an integrated diffusion component composed of plasterboard and a high-density wooden board, an air pressure balance device, a radon accumulation chamber, and a support structure. The uniformity and stability of the reference device were evaluated using the activated carbon- γ spectrometry method and the open-loop method, respectively, to measure the radon exhalation rate. The reference device achieved different radon exhalation rates by employing diffusion solid radon sources with different activities. Nineteen measurement points were uniformly selected on the radon exhalation surface of the reference device, and the uniformity of the radon exhalation rate was better than 5%. The short-term stability of the reference device was better than 5% under different environmental conditions and was almost unaffected by ambient air pressure, environmental temperature, and relative humidity.

Full Text

Preamble

Reference Device for Calibration of Radon Exhalation Rate Measuring Instruments and Its Performance

Weigang Li¹, Detao Xiao¹, Zhengzhong He¹, Xiangyuan Deng¹, Shoukang Qiu¹
¹*School of Nuclear Science and Technology, Hunan Provincial Key Laboratory of Radon, University of South China, Hengyang, Hunan Province 421001, China*
xiaodt@usc.edu.cn* (Corresponding author)

Abstract: Environmental radon emanates from the exhalation and release of soil, rocks, and building materials. Environmental radon contamination tracing and radon pollution prevention and control require the measurement of the radon exhalation rate on media surfaces. Reliable measurements of the radon exhalation rate cannot be achieved without regular calibration of the measuring instrument with a high-performance reference device. In this study, a reference device for the calibration of radon exhalation rate measuring instruments was developed using a diffusion solid radon source with a high and stable radon emanation coefficient, an integrated diffusion component composed of a plasterboard and a high-density wooden board, an air pressure balance device, a radon accumulation chamber, and a support structure. The uniformity and stability of the reference device were evaluated using the activated carbon- γ spectrum and open-loop method, respectively, to measure the radon exhalation rate. The reference device achieved different radon exhalation rates by using different activities of diffusion solid radon sources. Nineteen measurement points were regularly selected on the radon exhalation surface of the reference device, and the uniformity of the radon exhalation rate exceeded 5%. The short-term stability of the reference device was better than 5% under different environmental conditions and was almost unaffected by the ambient air pressure, environmental temperature, and relative humidity.

Keywords: radon exhalation rate; stability; uniformity; high-density wooden board; plasterboard

1 Introduction

Radon, a naturally occurring radioactive gas produced by the radioactive decay of ^{226}Ra [?], is the second leading cause of lung cancer after smoking [?] and was listed among the 19 most carcinogenic substances by the World Health Organization in 2009 [?]. Researchers have become increasingly interested in environmental radon exhalation rates (RERs) from media surfaces [?].

The RER is the radioactivity of radon exhaled from a media surface per unit area and time. In evaluating the efficacy of anti-radon coverage treatment in decommissioned uranium mines and metallurgy facilities, such as waste rock yards and tailings reservoirs, RER measurements are essential [?][?]. Additionally, source term inversion and radiation field reconstruction are crucial in alleviating environmental radon pollution [?]. Accurate RER measurements help determine the contribution of radon release to air pollution [?]. China has issued a standard (GB 50325-2020) for indoor environmental pollution control in civil building engineering [?], requiring that the measurement of the RER of building materials and the soil surface be performed continuously for 10 h using a monitor with a

detection limit of less than $1 \text{ mBq} \cdot \text{m}^{-2} \cdot \text{s}^{-1}$. Thus, an RER reference device should be developed to achieve reliable measurements of radon exhalation rates.

Researchers have been endeavoring to develop an RER reference device that achieves stability and uniformity and is almost unaffected by ambient air pressure, environmental temperature, and environmental relative humidity. Many RER measurement methods and devices have been developed [?][?][?][?]. However, the RER results measured by different monitors on the surface of the same media have significantly differed, and the performance parameters have been inconsistent. In 1983, in the United States, Artley et al. [?] used a Bendix Company calibration device to track thin-source radon exhalation. The instrument verification efficiency of the RER was determined by deriving the tailing ejection coefficient and medium porosity from the theoretical design of the device. According to the verification results, the efficiency was higher than 100%. In 1990, the National Institute of Standards and Technology attempted to develop a standard device for measuring the RER. Colle et al. [?] developed a device composed of an electroplated solid radon source and plastic film that did not achieve a stable RER. In 1999, the University of South China developed a verification device for RER measuring instruments with satisfactory stability and uniformity [?]. The stability of the device used since 2000 was found to be significantly affected by changes in ambient air pressure. In 2015, Lv Lidan [?] measured the RER on the surface of a device by using the activated carbon- γ spectrum method. The RERs ranged from 1.34 to $2.03 \text{ Bq} \cdot \text{m}^{-2} \cdot \text{s}^{-1}$, and the uniformity was higher than 20% at 21 measurement points. In 2016, Tsapalov established a standard device that operates under ideal (laboratory environment) and nonideal (outdoor) conditions, providing a standard model for the nonideal case [?]. However, the observed and theoretical data significantly differed when the model was used for testing. Thus, increasing the accuracy of calibration devices is necessary to calibrate the various RER measuring instruments on the market and assess their performance parameters.

This study aimed to develop an RER reference device that exhibits short-term RER stability and uniformity better than 5%, rapidly adjusts the RER at different levels, and satisfies the calibration requirements of various instruments.

2.1 Architecture of RER Reference Device (RERRD)

Ambient air pressure influences radon migration, and the pressure gradient causes seepage [?]. Thus, dense and low-permeability double-layer plate materials were used as the diffusion media for the reference device, and an air pressure balance device was designed to reduce the influence of ambient air pressure.

As shown in Fig. 1 [Figure 1: see original paper], the RERRD comprises five parts. A 0.6 L cylindrical solid radon source chamber was located at the bottom of the RERRD. A solid radon source with a high emissivity coefficient, manufactured by the University of South China, was sealed inside a solid radon source

chamber. This solid radon source had an emanation coefficient of $96 \pm 2\%$. A stainless steel sheet was stamped and welded into one piece to form the shell of the device, with a total height of 0.6 m and a surface smoothness of less than 2 mm. The manufacturing company tested the airtightness of the device during factory delivery, and the device passed the test with satisfactory airtightness.

The radon accumulation chamber was formed as an inverted cone, and the diffusion media (plasterboard and high-density wooden board) were shaped into cylinders to guarantee a one-dimensional upward direction of radon diffusion. The plasterboard, created by an artificial combination of solidification and drying, had well-developed, unevenly distributed internal pores [?]. The high-density wooden board was fabricated using the high pressure of machines, and small and regular interior pores were formed. There is a crucial small pore between the plasterboard and the high-density wooden board. After the radon diffused through the plasterboard, it was evenly distributed in the pores. The radon evenly distributed in the pores diffuses to the surface through the high-density wooden board and finally forms a uniform RER. The plasterboard and high-density wooden board were combined at a specific thickness to ensure the stability and uniformity of the RER on the media surface. Surface moisture has a significant impact on the uniformity of the RERRD. The emulsion paint exhibited excellent waterproofing ability. The high-density wooden board coated with the paint emulsion was essentially moisture-free. The radon in the accumulation chamber was highly concentrated, and a concentration gradient existed between the upper and lower surfaces of the diffusion medium. Therefore, the radon exchange in the media occurs via active diffusion.

2.2 Principle of Radon Diffusion in a Two-Layered Media

Radon exists as a single-atom gas at normal environmental temperatures, and its migration occurs via atomic thermal motion and convection. Because of the porous nature of the media, radon atoms can move freely through the connecting pores inside the media and diffuse out of the surface. The diffusion motion of radon in porous media obeys Fick's first law [?] is as follows:

where J is the RER on the surface of the RERRD ($\text{Bq} \cdot \text{m}^{-2} \cdot \text{s}^{-1}$), D is the radon diffusion coefficient in the media ($\text{cm}^2 \cdot \text{s}^{-1}$), and C is the radon concentration gradient in the media.

The second-order differential equation [?] for the diffusion and migration of radon through the medium is expressed by Equation (2). where x denotes the coordinate of the axis normal to the surface of the medium (m). $2 - (,) (,)$ The radon diffusion coefficient expresses the degree of radon diffusion in a medium. The relationship between the RER of the medium and the diffusion coefficient [?] is expressed by Equation (3).

When there is no pressure gradient in the surrounding space, the radon transport within the = - double-layer media is purely diffusion transport. Based on radon diffusion theories, when the radon concentration reaches a stable equilib-

rium state in space, the differential equation for radon diffusion in double-layer media is $D_1 \frac{dC_1}{dx} - \lambda C_1 = 0$ and $D_2 \frac{dC_2}{dx} - \lambda C_2 = 0$ where C_1 is the radon concentration in the plasterboard ($\text{Bq} \cdot \text{m}^{-3}$), C_2 is the radon concentration in the high-density wooden board ($\text{Bq} \cdot \text{m}^{-3}$), D_1 is the radon diffusion coefficient of the plasterboard ($\text{cm}^2 \cdot \text{s}^{-1}$), D_2 is the radon diffusion coefficient of the high-density wooden board ($\text{cm}^2 \cdot \text{s}^{-1}$), and λ is the radon decay constant (s^{-1}).

The ratio (k) between the RER on the lower surface of the plasterboard and that on the upper surface of the high-density wooden board was obtained as follows: where is the RER on the lower surface of the RERRD plasterboard, and upper surface of the high-density wooden board.

2.3.1 Instruments Selection

Various instruments were used in the experiments. The radon monitor used was an FYCDY-P30 produced by Hubei Fangyuan Environmental Protection Technology Co., Ltd. The constant air pump was manufactured by Schauenburg Technology. The radon collection hood, an integrated and forged cylindrical stainless steel plate, had a steel plate thickness of 2 mm, an inner diameter of 280 mm, a capacity of 5.85 L, and a bottom area of 0.06 m^2 (Fig. 2 [Figure 2: see original paper]). The activated carbon boxes, made by the Beijing Institute of Metrology, were flat aluminum cylindrical boxes containing 80 g activated carbon. The thickness of the carbon layer was 2.5 cm and was covered with a metal screen and sealing cap. The sampling area of the activated carbon boxes was $3.848 \times 10^{-3} \text{ m}^2$. The high-purity germanium gamma spectrometer was manufactured by the company ORTEC.

2.3.2 Open-Loop Method for Measuring RER

The open-loop method was used to measure the RER [?], in which the radon collection hood was inverted on the surface of the exhalation media. The radon generated in the radon collection hood was pumped into the environment using the constant air pump. We wanted to ensure that the amount of radon pumped into the environment was equal to that generated from the surface of the media; thus, radon-free air was simultaneously added to the radon collection hood, which was connected to the FYCDY-P30 radon monitor. Radon-free air is produced by absorbing radon from the air through an activated carbon canister. After the RERRD reached an equilibrium state, the radon concentration was obtained from the alpha particle counts produced by the radon progeny decay electrostatically collected on the surface radon detector [?]. The variation with time of the radon concentration in the radon collection hood was expressed using Equation (6).

When the radon concentration in the radon collection hood reaches a steady state, the left side $-\lambda C_1 - \lambda C_2 - 0$ of Equation (6) is 0.

Then, Because the natural decay constant of radon is $2.1 \times 10^{-6} \text{ s}^{-1}$, ($V \cdot$

C/S in Equation (7) can be neglected, $C_1 - C_0$ and Equation (7) can be simplified to $J = \frac{L(C_1 - C_0)}{S}$ where J is the RER to be measured on the surface of the exhalation media ($\text{Bq} \cdot \text{m}^{-2} \cdot \text{s}^{-1}$), L is the flow rate of the constant air pump ($\text{L} \cdot \text{min}^{-1}$), C_1 is the radon concentration in the radon collection hood ($\text{Bq} \cdot \text{m}^{-3}$), C_0 is the radon concentration in the environment ($\text{Bq} \cdot \text{m}^{-3}$), V is the volume of the radon collection hood (m^3), and S is the effective exhalation area of the radon collection hood (m^2).

2.3.3 Activated Carbon- γ Spectrum Method for Measuring the RER

The activated carbon- γ spectrum method [?] was used to measure the RER from the surface of the reference device. The activated carbon boxes were placed upside down on the surface of the exhalation medium to collect the radon exhaled by the RERRD during sampling. A high-purity germanium gamma spectrometer was used to measure the intensities of the characteristic peaks of the radon daughters in the activated carbon. Finally, the activated carbon- γ spectrum method was used to obtain the RER. The selected coconut shell-activated carbon had significant radon adsorption properties. Therefore, the impact of leakage and counter-diffusion was minimal [?].

First, the activated carbon boxes were sealed immediately after sampling. In this case, the short-lived daughters of radon (^{214}Pb and ^{214}Bi) reached radioactive equilibrium with ^{222}Rn [?]. This study used a high-purity germanium gamma spectrometer to measure the full energy peak area of the distinctive rays of the radon daughters; subsequently, the RER was calculated [?]. The 609 keV γ -rays from the decay of ^{214}Bi exhibited the highest branching ratio, the best energy peak shape, and least interference from other nuclides; thus, 609 keV γ -rays are generally used for experimental measurements [?].

To compare the results of multiple measurements, we selected a fixed area to measure the RER on the RERRD surface. Nineteen activated carbon boxes were evenly arranged in concentric circles on the surface of the exhalation media. Furthermore, these activated carbon boxes were used for measurements and distributed uniformly over the surface of the exhaled media. A schematic of the points on the activated carbon boxes is shown in Fig. 3 [Figure 3: see original paper]. The activated carbon boxes were dried at 120 °C for 8 h before the experiment. The background counting rate of the activated carbon boxes was measured after cooling, and the boxes were inverted equally on the surface of the exhaling media. The borders of the activated carbon boxes were sealed with plastic and stored for 24 h. The high-purity germanium gamma spectrometer was used to measure the net count rate after 3 h. The RERs were measured three consecutive times and calculated using Equation (9). where J is the RER on the surface of the media to be measured ($\text{Bq} \cdot \text{m}^{-2} \cdot \text{s}^{-1}$), K is the scale factor of () (1 - the high-purity germanium gamma spectrometer for 609 keV energy gamma rays (the scale factor was applied to the standard source scale before use) ($0.007 \text{ cps Bq}^{-1}$), n_r is the full energy

peak area (net count rates) of the characteristic 609 keV gamma rays (cps), n_b is the background counting rate of the activated carbon before the sampling rate (cps), S is the bottom area of the activated carbon boxes (m^2), TS is the sampling time (h), λ is the decay constant of radon (s^{-1}), and T is the time interval between stopping sampling of the activated carbon boxes and the start of the high-purity germanium gamma spectrometer measurement (h).

The uncertainty of the gamma spectrometry of the activated carbon- γ spectrum method mainly originates from the detection efficiency of the high-purity germanium gamma spectrometer and the uncertainty in the production process of the standard source. The high-purity germanium gamma spectrometer was calibrated using a 230 Bq standard source before use. The relative uncertainty of the spectrometer was approximately 5%, and the uncertainty in the production process of the standard source was approximately 3%. Because of the uncertainty of the experiment, the overall uncertainty of this method can be controlled within 10% [?].

2.4 Methods for Measuring the Media Radon Diffusion Coefficient

A miniaturized reference device was used to measure the radon diffusion coefficient of the medium (Fig. 4 [Figure 4: see original paper]). The type of diffusion medium could be changed arbitrarily. The volume of the radon accumulation chamber was V_1 (m^3); the radon concentration in the radon accumulation chamber was C_1 ($Bq \cdot m^{-3}$); and the change in radon concentration exhaled from the media was measured continuously using the open-loop method. The volume of the radon collection hood was denoted as V_2 (m^3); the radon concentration in the radon collection hood was denoted as C_2 ($Bq \cdot m^{-3}$); the bottom area of the radon collection hood was denoted as S (m^2); and the thickness of the media was denoted as H (m).

Then, $1 + 2(-1)$ When D/H^2 is sufficiently large, Equation (10) can be simplified to $1 =$ The radon diffusion coefficients of the plasterboard and high-density wooden board can be $1 \quad 2 \quad 1$ calculated using Equation (11).

2.5 Method for Rapidly Establishing Equilibrium Radon Concentration in Radon Accumulation Chamber

Radon in the accumulation chamber was generated from solid radon sources, which required an excessively long time to reach equilibrium. As shown in Fig. 1, to rapidly establish the equilibrium radon concentration in the radon accumulation chamber, a method was adopted wherein the radon accumulation chamber was filled with a flowing gas–solid radon source in the initial state. The flowing gas–solid radon source must be emptied before use. The radon accumulation chamber was filled using an air pump after accumulation for a certain period. The variation in the radon concentration in the accumulation

chamber is expressed by Equation (12). where A_0 is the activity of the flowing gas solid radon source (Bq), V is the volume of the radon accumulation chamber (m^3), C is the radon concentration in the radon accumulation chamber ($Bq \cdot m^{-3}$), J_1 is the RER on the lower surface of the plasterboard ($Bq \cdot m^{-2} \cdot s^{-1}$), S is the effective exhalation area (m^2), k is the solid radon source jet coefficient, and λ is the radon decay constant (s^{-1}).

When the radon concentration reaches the equilibrium state, the left side of Equation (12) is equal to zero; thus, the initial radon concentration, denoted C_e , is obtained. As $J_1 = k \cdot J_2$. Equation (12) is reduced to According to Equation (13), when the activity of the solid radon source and the RER on the surface of the media are known, the initial radon concentration in the radon accumulation chamber can be obtained. The time required to attain equilibrium was reduced by the rapid filling of the radon accumulation chamber with a flowing gas solid radon source. Increasing the speed of the formation of a stable RER on the RERRD is feasible.

2.6.1 Uniformity Measurement

The consistency of the RER at any location on the surface of the exhalation medium is indicated by the uniformity of the RERRD. The RERRD produced different RERs in the area used for each sampling event. Uniformity of the RERRD is necessary to ensure the accuracy of instrument calibration. The uniformity of the RERRD was indicated by a relative standard deviation of the RER of over 19 points. A smaller relative standard deviation indicated better uniformity of the RERRD.

The average RER over multiple points is The uniformity of the RERRD is indicated by the relative standard deviation as follows: $\sigma_1 = \frac{1}{n} \sqrt{\sum_{i=1}^n (J_i - \bar{J})^2}$ where σ_1 indicates the uniformity of the RERRD, J_i indicates the RER measured at point i ($Bq \cdot m^{-2} \cdot s^{-1}$), n is the number of measurement points, and \bar{J} is the average RER over n points ($Bq \cdot m^{-2} \cdot s^{-1}$).

2.6.2 Short-Term Stability Measurement

The short-term stability of RERRD indicates a change in its RER over a continuous period. The relative standard deviation of the measured RER was calculated by holding the radon collection hood upside down at the center of the exhalation medium and measuring the RER several times using the open-loop method.

The average of the multiple RER measurements is The short-term stability of the RERRD is indicated by the relative standard deviation as follows: $\sigma_2 = \frac{1}{n} \sqrt{\sum_{i=1}^n (J_i - \bar{J})^2}$ where σ_2 denotes the short-term stability of the RERRD, J_i denotes the RER of the central point measurement ($Bq \cdot m^{-2} \cdot s^{-1}$), n denotes the number of measurements, and \bar{J} denotes the average RER over n measurements ($Bq \cdot m^{-2} \cdot s^{-1}$).

3.1 Relationship Between RERs on Upper and Lower Surfaces of Diffusion Component

The diffusion coefficient of radon indicates its ability to migrate in media. After the radon concentrations in the radon accumulation chamber and the radon collection hood reached a dynamic equilibrium state, the diffusion coefficients of the media were calculated using Equation (11). The radon diffusion coefficients of the plasterboard and high-density wooden board displayed in Table 1 are closely related to the RER on the upper and lower surfaces of the diffusion component.

Table 1: Radon diffusion coefficients for plasterboard and high-density wooden board

Media	Thickness H (m)	Diffusion coefficient D ($\text{m}^2 \cdot \text{s}^{-1}$)
Plasterboard		2.94×10^{-10}
High-density Wooden board		1.42×10^{-10}

As the diffusion coefficient decreased, the ability of radon to migrate in the medium weakened. As shown in Table 1, the diffusion ability of radon on the plasterboard was lower than that on the high-density wooden board. Therefore, a high-density wooden board was placed atop the plasterboard in the double-diffusion media in the design of the RERRD.

On the basis of the data in Table 1, we analyzed k . By substituting D_1 , D_2 , h_1 , and h_2 into Equation (18), we obtained $k = \frac{D_1 + D_2}{h_1 + h_2}$. The radon concentration exhaled from the RERRD on the surface of the high-density wooden board is much lower than that in the radon accumulation chamber, namely, $C_{h_1+h_2} / C_{h_1} < 0$; therefore, $k < 1$ and $J_1 < J_2$. The radon in the radon accumulation chamber was exhaled into the environment through the plasterboard and high-density wooden board. The RERs from the lower surface of the plasterboard and the upper surface of the high-density wooden board were approximately equal. In this case, a solid radon source with different activities and RER can be rapidly obtained based on the initial radon concentration in the radon accumulation chamber by using Equation (11). The time required for the RERRD to reach equilibrium can be reduced considerably from 28 d to approximately 2 d, which fulfills the requirements of experimental measurements rapidly.

3.2 Uniformity Measurement Result of the RER of the RERRD

The uniformity of the RER at any point on the surface of the diffusion medium is a crucial index for measuring the uniformity of the RERRD. The radon accumulation chamber of the RERRD was filled with two different solid radon sources with activities of 1.0×10^5 Bq and 2.0×10^5 Bq. The RER of the

RERRD was measured thrice using the activated carbon- γ spectrum method. The measurement results are presented in Fig. 5 [Figure 5: see original paper].

Fig. 5: (a) Uniformity of RER after placing a 1.0×10^5 Bq solid radon source; (b) uniformity of RER after placing a 2.0×10^5 Bq solid radon source.

The mean RERs at the 19 points were 197.78 ± 2.32 and 399.26 ± 5.99 $m^{-2} \cdot s^{-1}$, and their relative standard deviations were 1.17% and 1.50%, respectively. The different points exhibited a uniformity of greater than 5%. The RERs measured at the points of the inner circle (2–7) swung up and down below the average. However, the RERs measured at points outside the circle (8–19) were higher than the average. The low efficiency of the high-purity germanium gamma spectrometer and the statistical errors in the measurement cannot be ignored. Another factor to consider is the latex paint sprayed on the surface of the high-density wooden board to eliminate the effects of moisture. Manual spraying could not ensure consistency in the thickness of the latex paint layer at each spot. The points with a thin latex paint layer had higher RERs than the other points did.

3.3 Short-Term Stability Measurement Results of the RER of the RERRD

The short-term stability of the RER characterizes whether it is affected by changes in ambient air pressure, environmental temperature, relative humidity, and other conditions. The two solid radon sources with activities of 1.0×10^5 and 2.0×10^5 Bq, respectively, were placed in the radon accumulation chamber. The radon collection hood was sampled and measured in the middle of the medium surface by using the open-loop method. The effects of the ambient pressure, environmental temperature, and relative humidity on the RER were studied.

3.3.1 Stability of the RER of the RERRD Under Various Ambient Air Pressure

The temperature inside the room was controlled via air conditioning and maintained at 20 °C. The dehumidifier stabilized the relative humidity of the room at 40%. The hourly ambient air pressure and the RER were measured thrice using the open-loop method. The daily variation curves of ambient air pressure and the RER are shown in Fig. 6 [Figure 6: see original paper].

Fig. 6: Variation in ambient air pressure and RER with time. (a) Daily variation curves of ambient air pressure and RER after placing a solid radon source with an activity of 1.0×10^5 Bq; (b) Daily variation curves of ambient air pressure and RER when placing a solid radon source with an activity of 2.0×10^5 Bq.

The ambient air pressure changed daily, with high values between 9 and 11 am

and low values between 3 and 5 am [?]. By contrast, the RERs resulting from the solid radon sources of 1.0×10^5 Bq and 2.0×10^5 Bq on the surface of the media of the RERRD varied little within 1 d, with daily relative standard deviations of 2.54% and 3.65%, respectively. Therefore, the short-term stability of the RERRD was greater than 4%. The pressure balance device of the RERRD can reduce the effects of ambient air pressure on radon migration. Internal changes in the pressure in the radon accumulation chamber were maintained at all times. We calculated the correlation coefficient between the ambient air pressure and the radon exhalation rate and found that the correlation between these two factors was weak (Fig. 6 (a): 0.42; Fig. 6 (b): 0.35). The results revealed that the ambient air pressure had almost no effect on the RER of the RERRD.

3.3.2 Stability of the RER of the RERRD Under Various Environmental Temperatures

The environmental relative humidity was stabilized at 40% with a dehumidifier while imposing different environmental temperatures via air conditioning. As shown in Fig. 6, the influence of ambient air pressure on the RER, although nonsignificant, should be eliminated as much as possible. The open-loop method was used to conduct the experiment from 9:00 to 11:00 am daily. The relationship between the environmental temperature and the RER of the RERRD was measured thrice using the open-loop method (Fig. 7 [Figure 7: see original paper]).

Fig. 7: Variation in RER with environmental temperature. (a) Relationship between RER and environmental temperature after placing a solid radon source with an activity of 1.0×10^5 Bq; (b) Relationship between RER and environmental temperature after placing a solid radon source with an activity of 2.0×10^5 Bq.

Fig. 7 shows the relationship between the RER and environmental temperature at a relative humidity of 40% and essentially unchanged ambient air pressure. Because the adjustable range of the air conditioning is 16–30 °C, the temperature was set to 16, 20, 25, and 30 °C, respectively; for a solid radon source with an activity of 1.0×10^5 Bq, the short-term stabilities were 1.02%, 0.76%, 1.52%, and 2.08%, respectively; by placing a solid radon source with an activity of 2.0×10^5 Bq, the short-term stability was 2.56%, 1.90%, 2.49%, and 4.14%, respectively. The short-term stability of the RERRD was within 5%. The environmental temperature had little effect on the RER of the RERRD.

3.3.3 Stability of the RER of the RERRD Under Various Environmental Relative Humidity

First, the indoor air conditioning stabilized the temperature at 20 °C. Subsequently, a cloudy and rainy day (with high relative humidity in the environmental air) was selected. The relative humidity of the indoor air was controlled

using a dehumidifier. To consider the influence of ambient air pressure, we used the open-loop method to conduct the experiment from 9:00 to 11:00 am. The effect of the environmental relative humidity on the RER of the RERRD is presented in Fig. 8 [Figure 8: see original paper].

Fig. 8: (a) Relationship between RER and environmental relative humidity after placing a solid radon source with an activity of 1.0×10^5 Bq; (b) Relationship between RER and environmental relative humidity after placing a solid radon source with an activity of 2.0×10^5 Bq.

The environmental temperature was 20 °C, and the ambient air pressure was essentially unchanged. The relationship between the RER and environmental relative humidity is shown in Fig. 8. When placing a solid radon source with an activity of 1.0×10^5 Bq, the relative humidity was set to 40%, 50%, 60%, 70%, 80%, and 90%; the short-term stabilities were 1.97%, 1.78%, 2.54%, 2.40%, 2.37%, and 2.59%, respectively; for a solid radon source with an activity of 2.0×10^5 Bq, the short-term stabilities were 2.70%, 2.27%, 1.48%, 1.61%, 3.29% and 3.10%, respectively. The relative standard deviation and short-term stability of the RER of the RERRD were within 4% at $T = 20$ °C. The environmental relative humidity had little effect on the RER measurement, and it had almost no effect on subsequent measurements using the RERRD.

3.4 Relationship Between Activities of Diffusion Solid Radon Sources and the RER of the RERRD

The studied RERRD can create different levels of uniform and stable RERs on the surface of the media by changing the activities of the solid radon sources. The existing activities of the solid radon sources in the Radon Laboratory of the University of South China were 3.0×10^3 , 8.5×10^3 , 1.7×10^4 , 1.0×10^5 , and 2.0×10^5 Bq. They were placed separately in the accumulation chamber of the RERRD, and the RER was rapidly brought to a steady state by filling the accumulation chamber with a predetermined amount of radon. The activated carbon- γ spectrum method was used to measure the RER by placing solid radon sources with different activities at the midpoint of the medium. The results are shown in Fig. 9 [Figure 9: see original paper].

Fig. 9: Relationship between RERs and activities of solid radon sources.

As shown in Fig. 9, a linear fitting equation between the RERs of the RERRD and the activities of the solid radon sources was obtained by linear fitting of the data as follows:

Because the minimum activity of the solid radon source used in our experiment was 3000 Bq, $A_0 = 0.002 A - 5.854$ Equation (19) is valid when $A_0 \geq 3000$.

R^2 was 0.99. If the activity of a radon source placed in the RERRD is known, the RER of the RERRD can be rapidly obtained using Equation (19). This

finding provides a reference activity for a solid radon source for high RER levels and improves the experimental efficiency.

4 Conclusion

In this study, an integrated diffusion assembly was formed from a plasterboard and high-density wooden board using the solid radon sources of 3.0×10^3 , 8.5×10^3 , 1.7×10^4 , 1.0×10^5 , and 2.0×10^5 Bq, and an air pressure balancing device was developed to eliminate the influence of the ambient air pressure. The characteristics of RERRD were tested.

The uniformity of the RER on the surface of the media was less than 5% when solid radon sources with different activities were inserted into the RERRD, and the short-term stability was less than 5%. The RER of the RERRD was almost unaffected by daily variations in ambient air pressure, environmental temperature, and environmental relative humidity. Therefore, the errors induced by these three factors can be disregarded. For calibration, the RERRD was set up in a laboratory under normal environmental conditions. If different RERs are required for an experiment, the activity of the solid radon source to be placed in the RERRD can be obtained rapidly. Moreover, the RERRD, designed as a detachable structure, allows experimental studies of RERs to be conducted by changing the diffusion media of different materials.

With excellent performance and rapid realization at various levels of RERs, the RERRD has the potential to be developed into a new type of calibration device for RER measuring instruments.

Acknowledgments

We are particularly grateful to Drs. Shaobin Guan, Zongjie Zhou, Fenglin Li, and Tao Sun of the Chinese Nuclear Industry Remote Sensing Center for providing support under experimental conditions.

Funding: This study was supported by the National Natural Science Foundation of China (Grant No. 11875165).

Author Contributions: All authors contributed to the study conception and design. Material preparation, data collection and analysis were performed by Weigang Li, Detao Xiao, Zhengzhong He, Xiangyuan Deng and Shoukang Qiu. The first draft of the manuscript was written by Weigang Li and all authors commented on previous versions of the manuscript. All authors read and approved the final manuscript.

Data Availability Statement: The data that support the findings of this study are openly available in Science Data Bank at <https://doi.org/10.57760/sciencedb.j00186.00136> and <https://cstr.cn/31253.11.sciencedb.j00186.00136>.

References

- [1] N. Lavi, V. Steiner, Z.B. Alfassi, Measurement of radon emanation in construction materials. *Radiation measurements*. 44(4): 396-400 (2009). <https://doi.org/10.1016/j.radmeas.2009.04.009>
- [2] United Nations Scientific Committee on the Effects of Atomic Radiation. *Effects of Ionizing Radiation, United Nations Scientific Committee on the Effects of Atomic Radiation (UNSCEAR) 2006 Report, Volume I: Report to the General Assembly, Scientific Annexes A and B*. United Nations, 2008. <https://doi.org/10.18356/7fb405cb-en>
- [3] Ting, David S-K. WHO Handbook on Indoor Radon: A Public Health Perspective. *International Journal of Environmental Studies*. Vol.67(1): 100-102 (2010). <https://doi.org/10.1080/00207230903556771>
- [4] Y. Huo, P. Xu, C. Zhou, et al., Effect of environment humidity to radon measurement with SSNTD. *Nuclear Science and Techniques*. 20(4): 228-230 (2009). <https://doi.org/10.13538/j.1001-8042/nst.20.2>
- [5] J. Somlai, Z. Gorjánác, A. Várhegyi, et al., Radon concentration in houses over a closed Hungarian uranium mine. *Science of the total Environment*. 367(2-3): 653-665 (2006). <https://doi.org/10.1016/j.scitotenv.2006.02.043>
- [6] J. Jónás, J. Somlai, E. Tóth-Bodrogi, et al. Study of a remediated coal ash depository from a radiological perspective, *Journal of environmental radioactivity*. 173: 75-84 (2017). <https://doi.org/10.1016/j.jenvrad.2016.11.010>
- [7] X.J. Li, S.K. Qiu, C.K. Liu, A calibration facility for radon fluxmeter. *Radiation Protection (Taiyuan)*. 28(4): 197-201 (2008).
- [8] S. Röttger, A. Röttger, C. Grossi, et al., Radon metrology for use in climate change observation and radiation protection at the environmental level. *Advances in Geosciences*. 57: 37-47 (2022). <https://adgeo.copernicus.org/articles/57/37/2022/>
- [9] W.P. YUNFEI, Reduce indoor pollution and share fresh air –interpretation of GB 50325-2020 “Standard for door Environmental Pollution Control of Civil Building Engineering”. *Construction Quality*. 39(03): 1-5 (2021). (in Chinese)
- [10] L.D. Lv, Z.Z. He, S.K. Qiu, et al., Evaluation and measurement methods for the surface radon exhalation rate of buildings. *Indoor and Built Environment*. 31(10): 2378-2385 (2022). <https://doi.org/10.1177/1420326X221109754>
- [11] Y.F. Yang, L.D. Lv, S.K. Qiu, et al. Study on the influence of sampling methods for measuring soil radon exhalation rates, *Radiation Measurements*. 159: 106880 (2022). <https://doi.org/10.1016/j.radmeas.2022.106880>
- [12] GB/T(Chinese Standards) GB/T 16143-1995. Charcoal canister method for measuring ^{222}Rn exhalation rate from building surface (English Vision). Beijing: National Health Commission of the People’s Republic of China, 1995.
- [13] Y.L. Tan, D.T. Xiao, H. Yuan, et al., Revision for measuring radon exhalation rate in open loop. *Journal of Instrumentation*. 8(01): T01004 (2013). <https://doi.org/10.1088/1748-0221/8/01/T01004>
- [14] J.N. Hartley, G.W. Gee, E.G. Baker, et al., 1981 radon barrier field test at Grand Junction uranium mill tailings pile. Pacific Northwest Lab. Richland, WA (USA), 1983. <https://www.osti.gov/biblio/6405>
- [15] R. Collé, J.M.R. Hutchinson, M.P. Unterweger, The NIST primary radon-

- 222 measurement system. *Journal of research of the National Institute of Standards and Technology*. 95(2): 155 (1990). <https://doi.org/10.6028/jres.095.018>
- [16] L.D. Lv, X.P. Qiu, S.K. Qiu, et al., Study of active carbon measuring method for accurate measurements of radon exhalation rates for building materials. *Journal of University of South China (Science and Technology)*. 28(01): 9-12 (2014). <https://doi.org/10.19431/j.cnki.1673-0062.2014.01.004> (in Chinese)
- [17] A. Tsapalov, K. Kovler, P. Miklyav, Open charcoal chamber method for mass measurements of radon exhalation rate from soil surface. *Journal of Environmental Radioactivity*. 160: p. 28-35 (2016). <https://doi.org/10.1016/j.jenvrad.2016.04.016>
- [18] X.S. Liu, S.K. Qiu, A rapid and accurate method for measuring radon exhalation rate. *Radiation Protection*. 27(03): 156-162 (2007). <https://doi.org/10.3321/j.issn:1000-8187.2007.03.005> (in Chinese)
- [19] Q. Tang, S.K. Qiu, D.T. Xiao, et al., Extraction and purification of ^{227}Ac and development of solid ^{219}Rn source. *Radiochimica Acta*. 102(1-2): 169-174 (2014). <https://doi.org/10.1515/ract-2014-2117>
- [20] J. Ke, Y.L. Ke, Y.H. Zhuang, et al., The Analysis of the Performance in the Calibration of Radon Measurement Instrument by Several Kinds of Solid Standard Source. *South China Journal of Seismology*. 35(2): 43-49 (2015). <https://doi.org/10.13512/j.hndz.2015.02.007> (in Chinese)
- [21] W.Y. Du, GD—L2 Flow-type Solid radon Source Calibration Factors of Radon Measurement Instrument and Its Influence. *J. of Institute of Disaster Prevention*. 15(3): 67-71 (2013). <https://doi.org/10.3969/j.issn.1673-8047.2013.03.012> (in Chinese)
- [22] E. Saleh, A.M.A. Al-Sobahi, S.A.E. El-Fiki, Assessment of radon exhalation rate, radon concentration and annual effective dose of some building materials samples used in Yemen. *Acta Geophysica*. 69(4): 1325-1333 (2021). <https://doi.org/10.1007/s11600-021-00628-z>
- [23] M. Noguchi, H. Wakita, A method for continuous measurement of radon in groundwater for earthquake prediction. *Journal of Geophysical Research*. 82(8): 1353-1357 (1977). <https://doi.org/10.1029/JB082i008p01353>
- [24] H.N.P. Thu, N.Van Thang, L.C. Hao, The effects of some soil characteristics on radon emanation and diffusion. *Journal of Environmental Radioactivity*. 216(C): 106189 (2020). <https://doi.org/10.1016/j.jenvrad.2020.106189>
- [25] C. Miró, E. Andrade, M. Reis, et al., Development of a couple of methods for measuring radon exhalation from building materials commonly used in the Iberian Peninsula. *Radiation protection dosimetry*. 160(1-3): 177-180 (2014). <https://doi.org/10.1093/rpd/ncu063>
- [26] Z.Q. Li, D.T. Xiao, G.Z. Zhao, et al., Rapid determination of radon monitor's calibration factors. *Nuclear Science and Techniques*. 27(5): 1-5 (2016). <https://doi.org/10.1007/s41365-016-0118-2>
- [27] P. Tuccimei, M. Moroni, D. Norcia, Simultaneous determination of ^{222}Rn and ^{220}Rn exhalation rates from building materials used in Central Italy with accumulation chambers and a continuous solid state alpha detector: influence of particle size, humidity and precursors concentration. *Applied radiation and isotopes*. 64(2): 254-263 (2006). <https://doi.org/10.1016/j.apradiso.2005.07.016>

- [28] G.P. Xiao, D.T. Xiao, L.D. Lv, et al., Study on active carbon sampling device for measuring radon exhalation rate. *Nuclear Electronics and Detection Technology*. 35(11): 1119-1123 (2015). <https://doi.org/10.3969/j.issn.0258-0934.2015.11.016> (in Chinese)
- [29] C.E. Lawrence, R.A. Akber, A. Bollhöfer, et al, Radon-222 exhalation from open ground on and around a uranium mine in the wet-dry tropics. *Journal of Environmental Radioactivity*. 100(1): 1-8 (2009). <https://doi.org/10.1016/j.jenvrad.2008.09.003>
- [30] S.H. Alharbi, R.A. Akber, Radon-222 activity flux measurement using activated charcoal canisters: Revisiting the methodology. *Journal of environmental radioactivity*. 129: 94-99 (2014). <https://doi.org/10.1016/j.jenvrad.2013.12.021>
- [31] J.L. Pinault, J.C. Baubron, Signal processing of diurnal and semidiurnal variations in radon and atmospheric pressure: a new tool for accurate in situ measurement of soil gas velocity, pressure gradient, and tortuosity. *Journal of Geophysical Research: Solid Earth*. 102(B8): 18101-18120 (1997). <https://doi.org/10.1029/97JB00971>

Note: Figure translations are in progress. See original paper for figures.

Source: ChinaXiv — Machine translation. Verify with original.